

Tel: Email: Web:

+001.301.228.2444 info@xpressbio.com www.xpressbio.com

## **Certificate of Analysis** Mouse Adenovirus-2 (K87) **PCR Positive Control**

Catalog Number: PCRPC130

Date: October 2018

-XpressBio

**Volume**: 250 µl, 250 Reactions

**DNA Concentration:** 200 ng/μl of DNA in TE buffer. Use 0.5 to 1.0 μl per PCR reaction.

**Specific Activity:** At least 500 copies per µl.

Storage conditions: -20° C to -70° C

Typical PCR reaction: Total volume 50 μl, 25 μl water, 5 μl 10x PCR Buffer, 2 μl 50

mM MgSO<sub>4</sub>, 1 μl 10 mM dNTP mix, 1 μl 10 μM Forward Primer, 1 μl 10 μM Reverse Primer, DNA 0.5 to 1 µl, add water to 49.5 µl, and add last 0.5 µl Taq Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

## **Typical Thermal Cycler Program:**

1.	94°C	2 min
2.	94°C	15 sec
3.	55°C	30 sec
4.	0.5°/sec	to 68°C
5.	68°C	2 min
6.	Go to 2	35 times
7.	72°C	5 min
8.	4°C	forever
9.	End	

PCR Primers: FOR 5'- AAG CGC ACC TAC GAT TAC AT-3'

Hexon 252 bases

REV 5'-CTT CCT GAA AGC CCA CTC-3'

References: Grove, KA, Smith, PC, Booth, CJ, and Compton, SR, J Am Assoc Lab Anim

Sci 51: 789-796, 2012.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.