

Certificate of Analysis **Mycoplasma Detection Kit**

MycoXpress Mycoplasma PCR Detection Kit

Lot Number:

Catalog Number:

CSD100

Date of Analysis:

2012. 02. 03

Test:

The lot has been tested for the ability to amplify DNA from *Mycoplasma fermentans*.

Acceptance Criteria and Certified Values:

Detection limit of 150 genome copies per 100 μ l of culture supernatant that corresponds to 3 mycoplasma and no cross-reactivity amplification. Reactions without template must be negative.

This certificate is a declaration of analysis at the time of manufacture and does not release you from performing the standard control upon receipt of incoming goods. The manufactured product satisfies with all forced criteria. The lot was released for sale.