



+001.301.271.5445 Email: info@expressbiotech.com Web: www.expressbiotech.com

Tel:

## **Certificate of Analysis**

## Streptobacillus moniliformis **PCR Positive Control**

Catalog Number: PCRPC183

Date: April 2016

Volume: 250 µl, 250 PC Reactions

**DNA Concentration:** 200 ng/μl of DNA in TE buffer. Use 0.5 to 1.0 μl per PCR reaction.

**Specific Activity:** At least 500 copies per µl.

Storage conditions: -20° C to -70° C

Typical PCR reaction: Total volume 50 µl, 25 µl water, 5 µl 10x PCR Buffer, 2 µl 50

mM MgSO<sub>4</sub>, 1 μl 10 mM dNTP mix, 1 μl 10 μM Forward Primer, 1 μl 10 μM Reverse Primer, DNA 0.5 to 1 µl, add water to 49.5 µl, and add last 0.5 µl Taq Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

## **Typical Thermal Cycler Program:**

1.	94°C	2 min
2.	94°C	15 sec
3.	55°C	30 sec
4.	0.5°/sec	to 68°C
5.	68°C	2 min
6.	Go to 2	35 times
7.	72°C	5 min
8.	4°C	forever
9.	End	

PCR Primers: FOR 5'-CAT ACT CGG AAT AAG ATG G-3' PCR product is 269 bases REV 5'-GCT TAG CTC CTC TTT GTA C-3'

**References:** Kimura, M. et al., Detection of Streptobacillus spp. in feral rats by specific polymerase chain reaction. Microbiol Immunol 52:9-15, 2008.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.