

Certificate of Analysis

Mouse Hepatitis Virus

PCR Positive Control

Date: May 24, 2011

Catalog Number: PCRPC105

Volume: 250µl, about 250 reactions

RNA Concentration: 200 ng/µl in TE buffer. Use 0.5 to 1.0 µl per RT-PCR reaction.

Storage conditions: -20° C to -70° C

Expiration: 3 years from date of shipment

First Strand Reaction: primed with oligo dT or anti-sense REV oligo at 1 µl of a 1 µM solution.

Typical PCR reaction: Total volume 50 µl, 25 µl water, 5 µl 10x PCR Buffer, 2 µl 50

mM MgSO₄, 1 µl 10 mM dNTP mix, 1 µl 10 µM Forward Primer, 1 µl 10 µM Reverse Primer, DNA 0.5 to 1 µl, add water to 49.5 µl, and add last 0.5 µl *Taq* Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

Typical Thermal Cycler Program:

1. 94°C; 2 min
2. 94°C; 15 sec
3. 55°C; 30 sec
4. 0.5°/sec; to 68°C
5. 68°C; 2 min
6. Go to 2; 35 min
7. 72°C; 5 min
8. 4°C; forever
9. End

PCR Primers: FOR 5'-GTC ATG AGG CTA TTC CTA CTA-3'

REV 5'-ATA CAC ATC TTT GGT GGG-3'

Anti-sense REV 5'-CCC ACC AAA GAT GTG TAT-3'

References: Comparative Medicine **54:4**, August 2004, page 384.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.