

Certificate of Analysis

Rat Parvovirus (RPV)

PCR Positive Control DNA

Date: May 24, 2011

Catalog Number: PCRPC133

Volume: 250µl, about 250 reactions

DNA Concentration: 150 ng/µl in TE buffer. Use 0.5 to 1.0 µl per PCR reaction.

Storage conditions: -20° C to -70° C

Expiration: 3 years from date of shipment

Typical PCR reaction: Total volume 50 µl, 25 µl water, 5 µl 10x PCR Buffer, 2 µl 50

mM MgSO₄, 1 µl 10 mM dNTP mix, 1 µl 10 µM Forward Primer, 1 µl 10 µM Reverse Primer, DNA 0.5 to 1 µl, add water to 49.5 µl, and add last 0.5 µl *Taq* Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

Typical Thermal Cycler Program:	1. 94° C	2 min
	2. 94° C	15 sec
	3. 55° C	30 sec
	4. 0.5°/sec	to 68° C
	5. 68° C	2 min
	6. Go to 2	35 times
	7. 72° C	5 min
	8. 4° C	forever
	9. End	

PCR Primers: For 5'-CGC ACA TGT AGA ATT TTT GCT G-3'

Rev 5'-CAA AGT CAC CAG GCA ATG TGT T-3'

References: RPV-1 and RMV-1 PCR Assays, *Laboratory Animals* 2006. **40**: 64. and Wan, C., et, al., *J. General Virology* 2002. **83**: 2075-2083.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.