

Certificate of Analysis

Helicobacter bilis

PCR Positive Control

Date: January 2012

Catalog Number: PCRPC300

Volume: 250µl, about 250 reactions

DNA Concentration: 100 ng/µl in TE buffer. Use 0.5 to 2.0 µl per PCR reaction.

Storage conditions: -20° C to -70° C

Expiration: 3 years from date of shipment

Typical PCR reaction: Total volume 50 µl, 25 µl water, 5 µl 10x PCR Buffer, 2 µl 50

mM MgSO₄, 1 µl 10 mM dNTP mix, 1 µl 10 µM Forward Primer, 1 µl 10 µM Reverse Primer, DNA 0.5 to 1 µl, add water to 49.5 µl, and add last 0.5 µl *Taq* Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

Typical Thermal Cycler Program:

1. 94°C; 2 min
2. 94°C; 15 sec
3. 55°C; 30 sec
4. 0.5°/sec; to 68°C
5. 68°C; 2 min
6. Go to 2; 35 min
7. 72°C; 5 min
8. 4°C; forever
9. End

PCR Primers: FOR 5-ATG GAA CAG ATA AAG ATT TTA AAG CAA CTT CAG-3'

REV 5'-CTA TGC AAG TTG TGC GTT AAG CAT-3'

References: Comparative Medicine **54:4**, August 2004, page 384. and Feng, S. et. al.

Clinical and Diagnostic Laboratory **12:4**, pages 531-536, 2005.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.