



Express Biotech International
4650 Wedgewood Blvd
Suite 103
Frederick, MD 21703

Tel: +001.301.228.2444
Email: info@xpressbio.com
Web: www.xpressbio.com

Certificate of Analysis

Toolan's H1

PCR Positive Control

Catalog Number: PCRPC114

Date: August 2012

Volume: 250 µl

DNA Concentration: 150 ng/µl in TE buffer. Use 0.5 to 1.0 µl per PCR reaction.

Specific Activity: At least 500 copies per µl.

Storage conditions: -20° C to -70° C

Typical PCR reaction: Total volume 50 µl, 25 µl water, 5 µl 10x PCR Buffer, 2 µl 50

mM MgSO₄, 1 µl 10 mM dNTP mix, 1 µl 10 µM Forward Primer, 1 µl 10 µM Reverse Primer, DNA 0.5 to 1 µl, add water to 49.5 µl, and add last 0.5 µl *Taq* Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

Typical Thermal Cycler Program:

- | | | |
|----|----------|----------|
| 1. | 94°C | 2 min |
| 2. | 94°C | 15 sec |
| 3. | 55°C | 30 sec |
| 4. | 0.5°/sec | to 68°C |
| 5. | 68°C | 2 min |
| 6. | Go to 2 | 35 times |
| 7. | 72°C | 5 min |
| 8. | 4°C | forever |
| 9. | End | |

PCR Primers: FOR 5'-CTA GCA ACT CTG CTG AAG GAA CTC-3'

REV 5'-TAG TGA TGC TGT TGC TGT ATC TGA G-3'

References: Besselsen, DG, et.al. J. Clinical Microbiology **33:7** July 1995, pages 1699-1703.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.