

Tel: +001.301.228.2444 Email: info@xpressbio.com Website: www.xpressbio.com

Certificate of Analysis Mouse Pasteurella haemolytica

PCR Positive Control

Catalog Number: PCRPC144

Date: October 2013

Volume: 50 µl, 50 PC Reactions

DNA Concentration: 150 ng/μl in TE buffer. Use 0.5 to 1.0 μl per PCR reaction.

Specific Activity: At least 500 copies per µl.

Storage conditions: -20° C to -70° C

Typical PCR reaction: Total volume 50 μl, 25 μl water, 5 μl 10x PCR Buffer, 2 μl 50

mM MgSO₄, 1 μ l 10 mM dNTP mix, 1 μ l 10 μ M Forward Primer, 1 μ l 10 μ M Reverse Primer, DNA 0.5 to 1 μ l, add water to 49.5 μ l, and add last 0.5 μ l Taq Polymerase. Mix, spin down gently, and put tubes in thermal cycler. Alternatively, a master mix including everything except some water and DNA may be used.

Typical Thermal Cycler Program:

1.	94°C	2 min
2.	94°C	15 sec
3.	55°C	30 sec
4.	0.5°/sec	to 68°C
5.	68°C	2 min
6.	Go to 2	35 times
7.	72°C	5 min
8.	4°C	forever
9	End	

PCR Primers: FOR 5'-GCA GTG AAA GAA TTC TTT GGT TC-3'

REV 5'-GTT GCA TGT TAG AAC CCA T-3'

References: Dole, VS, Comparative Medicine **60:6** December 2010, pages 427-435.

This product is distributed for laboratory research use only. The safety and efficacy of this product in diagnostic or other clinical uses has not been established.